


SPB uotom ini 4/
4-aminofenol (itus (2)



BIN CARD PENGGUNAAN KOLOM

Form no. : RDC.2-FORM/022.00

Berlaku Tgl : 02 Desember 2024

Nama & Merk : Zorbax SB-C8, 250 x 4,6mm 5µm
Dimensi : L9
No. Batch/Seri/Lot : khusus PCT Infus
Tgl. Kedatangan : aktivasi, 13/01²⁵

C70

[illegible]

Cuci sebelum : 80 % MeOH
w % MeOH

Cuci : air : MeOH (940 : 60)
air (100 %)
MeOH : air (80 : 20)

C73

mersi		BIN CARD PENGGUNAAN KOLOM		Form no. RDC.2-FORM/022.00 Berlaku Tgl : 02 Desember 2024		
Nama & Merk		: Agilent LC Columns Zorbax SB-C8				
Dimensi		: 4.6 x 250 mm 5 micron				
No. Batch/Seri/Lot		: B24254 (C73)				
Tgl Kedatangan						
No	Tanggal	Analisa	Jumlah injek	Total injek	inisial	Keterangan
1.	16.01.25	Gabapentine				
2.	17.01.25	Aktivasi			WDP	100 % MeOH ; 80 % MeOH
3.	17.01.25	UOT Gabapentin pht 45	180	180	WDP	baik
4.	20.01.25	Cemaran olanzapin	50	230	HPH	baik
5.	20.01.25	Cemaran olanzapin	50	280	HPH	baik
6.	22.01.25	UOT Quetiapin	15	330 285	AHS	baik
7.	21.02.25	PK Quetiapine	40	335	AHS	baik
8.	10.02.25	Cemaran PCT Infus	110	445	ENA	baik
9.	03/05-03-25	Cemaran Cyano TMA	80	525	HPH FWH	baik
10.	10-12/03/25	UOT Quetiapine	47	572	WDP DYN	baik (peak belah)
14.	14.03.25	PCT Infus Cemaran	42	614	DYN	baik
15.	10.04.25	Fluoxetine HCl BB	27	621	ENA	mulai belah
16.	10.04.25	Fluoxetine HCl BB	80	701	ENA	Dibalik
17.	14.04.25	Fluoxetine HCl BB	3	704	CAG	searah jekt, dibalik mental
18.	15.04.25	Fluoxetine HCl BB	80	624 784	CAG	Di oven 100°, dibalik
19.	02.05.25	Fluoxetine HCl BB	162	946	HPH	dibalik
20.	20.05.25	TMA Cyanocobalamin	1	947	AMK	dibalik
21.	07.06.25	Cemaran olanzapin	26	973	FTA	Resolusi tak masuk
22.	25.06.25	Cemaran PCT Infus	92	1065	DYN	Tailing & plate count
23.	29.08.25	Elitac	26	1091	HPH	Tailing jekt