



PT. ABADINUSA USAHASEMESTA

Certificate

This is to certify that

Hanson - AUTOPLUS

Serial No.	: 1222-1531
Software Version	: -
Firmware Version	: -
Customer Name	: PT. MERSI FARMA TIRMAKU MERCUSANA
Address	: JL.Raya Bogor KM. 31 No.52, Cisalak, Depok


Has been tested and calibrated (the validation / calibration report / print out is attached), in accordance with the manufactures specifications

Test date :

25 AGUSTUS 2025

Performed by, **PT. AbadiNusa**

Engineer


AbadiNusa
Ananto Adic

User


Juwita Kristiani




TELEDYNE
HANSON RESEARCH

CALIBRATION REPORT HANSON - DISSOLUTION TESTER

Hanson	Model No.	Hanson Autoplus
	Drive Head No.	1222-1531

Customer	PT. Mersifarma Tirmaku Mercusana
Address	Jl. Raya Bogor No.52, Cisalak, Kota Depok

Density	0.998 g/mL
---------	------------

	Target	No.	Detected	Status
Hasil Nilai Low Volume	5 ml $\pm$ 0.05 mL	1	4,998	Passed
		2	5,009	Passed
		3	5,012	Passed
		4	5,011	Passed
		5	4,999	Passed
		6	4,999	Passed
Hasil Nilai High Volume	10 ml $\pm$ 0.10 mL	1	9,978	Passed
		2	9,982	Passed
		3	9,987	Passed
		4	9,983	Passed
		5	9,989	Passed
		6	9,983	Passed
Timer	5 minute $\pm$ 2 %		4 : 59 : 13	Passed
	10 minute $\pm$ 2 %		10 : 00 : 08	Passed

Tested by :

PT. AbadiNusa UsahaSemesta

AbadiNusa

(Ananto Adie)

Date : 25 Agustus 2025

End User


(Juwita Restiani)
