

ANALISA PERMINTAAN BARANG/JASA

SPB No : PF001988
Tanggal Pengajuan : 10 September 2026

No.	Deskripsi	Jumlah	Satuan	Analisa Kebutuhan	Due Date
1.	Biological Safety Cabinet Class IIA – Ductless, with external dimension NMT 150 x 85 x 180 cm. Equipped with HEPA 14 filter with efficiency 99.995% at 0.3 um. Material should be easy to clean and VHP resistant – With display, UV, light, sash, alarm, internal electric socket, and castor – Including training, IQ OQ, and calibration – Placement: Endotoxin room (211), Positive control room (207), and Microbial limit room (208)	3	Unit	Support microbial testing: Bioburden, Bacterial endotoxin, dan Specific bacteria pada sampel water system (PW dan WFI). Ke depannya selain untuk sampel water system dapat juga digunakan untuk sampel bahan baku dan kemas, IPC, serta produk jadi	9 Oct 2026

Dibuat oleh,



Diny Lucyanie

Diketahui oleh,



Elvina Fitrisia

Disetujui oleh,



Dyah Arum M.