

Bagian : QC

Due date : 5-Oct-26

Tanggal : 29-Sep-26

ANALISA PERMINTAAN BARANG

No	Kode Barang	Nama Barang	Satuan	Stok Saat Ini (A)	Rencana Kebutuhan (B)	Kekurangan (A -B + C)	Pending Penerimaan (C)	Jumlah	Realisasi Pembelian	Realisasi Penggunaan
1	ATSP01123	Dessicant cartridge with purge fitting	Un	0	1	-1	0	1	0	0
2	ATSP01122	Polystyrene	Pc	0	1	-1	0	1	0	0

Catatan :

1. Desiccant Cartridge with Purge Fitting, Part.no : G8043-67401


- Berfungsi mengurangi kelembapan (uap air) di dalam jalur optik/instrumen, melindungi komponen optik dari kelembapan yang dapat menyebabkan kondensasi atau dalam jangka panjang memengaruhi kondisi komponen.
- Bagian ini merupakan consumable part yang perlu dilakukan penggantian secara rutin (setahun sekali) agar kinerja untuk mengurangi kelembapan pada alat FTIR dapat bekerja dengan optimal
- Estimasi harga : Rp. 5.598.000/un
- Pemesanan pertama kali dept.QC

2. Polystyrene sample, Part.no : 2010128000


- Sebagai standar untuk kalibrasi dan verifikasi instrumen FTIR Agilent Cary 630
- Belum tersedia standar Polystyrene di QC, saat kalibrasi dipinjamkan oleh Supplier sehingga perlu dibuat pengadaan karena Polystyrene juga digunakan untuk verifikasi rutin yang direkomendasikan oleh teknisi.
- Estimasi harga : Rp. 3.476.000
- Pemesanan pertama kali dept.QC

Terlampir Service Report dan penawaran harga,

Dibuat


.....

Mengetahui


.....



QUOTATION

No. : 1638/Q/UAP/PART/IX/26
(please mention this number on your Purchase Order)

To : **PT Mersifarma**

Date : Sep 29, 2026

Dear Sir/Madam,

We are kindly submitting our quotation for :

No	Qty	Part Number	Description	Price/Qty (IDR)
1	1	2010128000	Polystyrene Sample	3.476.000,-

CONDITIONS :

- ❖ Price : Nett Franco Jakarta, Excluding VAT
- ❖ Del. Time : Ready Stock (stock is not binding)
- ❖ Validity : 2 weeks
- ❖ Term Of Payment : **14 days after received Invoice**
- ❖ Min Order : **Rp. 5.000.000,-**

We hope our quotation could meet your requirement and if you need any further information, please do not hesitate to contact us.

Best Regards,

*This document is generated by computer
No signature required*

Ingrit

Admin

Cc. Alsinus SR

Cc. Sheren

For more information, please contact:

Tlp : 021-53660641/42 (ext.311)

ingrit@unitamaanalitika.co.id

Please send your PO also CC to unitama@unitamaanalitika.co.id

Office:

Jl. Panjang No. 37C, Kebon Jeruk, Jakarta Barat 11530
Tlp. 021-53660641 / 53660642
Fax. 021-53661244

Ruko Puri Mas A2 No. 12, Jl. I Gusti Ngurah Rai
Gunung Anyar, Surabaya Timur 60294
Tlp. 031-87854672 / Fax. 031-87854793

Jl. Garuda No. 31C, Kemayoran, Jakarta Pusat 10610
Tlp. 021-4268478
Fax. 021-4268481



QUOTATION

No. : 1547/Q/UAP/PART/IX/26
(please mention this number on your Purchase Order)

To : PT. Mersifarma Sukabumi

Date : September 16, 2026

Dear Sir/Madam,

We are kindly submitting our quotation for :

No	Qty	Part Number	Description	Price/Qty (IDR)
1	1	G8043-67401	Dessicant cartridge with purge fitting	5.598.000,-

CONDITIONS :

- ❖ Price : Nett Franco Sukabumi, Excluding VAT, Excluding Installation
- ❖ Del. Time : Ready Stock (stock is not binding)
- ❖ Validity : 2 weeks
- ❖ Term Of Payment : Cash On Delivery/100% Cash in Advance

We hope our quotation could meet your requirement and if you need any further information, please do not hesitate to contact us.

Best Regards,

*This document is generated by computer
No signature required*

Sheren

Admin

Cc. Alsinus SR

Cc. Ingrit

For more information, please contact:

Tlp : 021-53660641/42 (ext.311)

sheren@unitamaanalitika.co.id

***Please send your PO also CC to unitama@unitamaanalitika.co.id**

Office:

Jl. Panjang No. 37C, Kebon Jeruk, Jakarta Barat 11530
Tlp. 021-53660641 / 53660642
Fax. 021-53661244

Ruko Puri Mas A2 No. 12, Jl. I Gusti Ngurah Rai
Gunung Anyar, Surabaya Timur 60294
Tlp. 031-87854672 / Fax. 031-87854793

Jl. Garuda No. 31C, Kemayoran, Jakarta Pusat 10610
Tlp. 021-4268478
Fax. 021-4268481

SERVICE REPORT

Customer	PT Mercifarma TPT	Instrument	Cary
Address	Subabumi	Model	630 FTIR
PIC		Serial No.	11/2331CU02
Email			
Job Title		PO / Wo No.	

Service Categories

☐ Installation	☐ Troubleshooting	☐ Training	☐ Repairing	☒ Maintenance
☐ Under Warranty	☐ Under service contract	☐ Emergency Call	☐ New Unit	☒ Calibration

Problems Reported by User :

Preventive Maintenance dan Calibration
--

Description of Work Performed :


Telah dilakukan PM terhadap alat Instrument Cary 630 FTIR meliputi:
1. Check Visual
2. Ganti Desiccant untuk tempat telah merupakan indikator rekomendasi Ganti
dan dokumentasikan penggantian Desiccant 1 Tahun 1kali
3. Lakukan Test STN
4. Lakukan Stability Test
5. Lakukan Test Wavenumber Accuracy
6. Lakukan Test Spectral Resolution
Dari pengujian diatas alat dalam kondisi baik dan dapat digunakan
Analisa
Note: User harus pemberian Polystyrene untuk kalibrasi selanjutnya
lingka untuk sebagai pemakai Polystyrene dari kantor Unitama

Recommendations :

* Polystyrene	PN: 2010128000
* Desiccant	11: 68093-67901

Working Hours :

Date	Time	Date	Time
23-5-2026			

Work completed on : Work performed by :  Rizal P Engineer	Checked by :  Service Manager	Customer Acceptance :  Customer Sign, Name and Stamp
---	---	--